

OW Series Digital Multimeter Technical Specifications

Standard conditions: The environment temperature is 18°C to 28°C, the relative humidity is less than 80%.

Note: When measuring AC voltage/current or capacitance, accuracy guarantee range is 5% to 100% of the range.

OW18D/OW18E multimeter

Function		Measurement Range	Resolution	Function		
DC Voltage (V)	mV ^[1]	20.000mV	0.001mV	$\pm(0.05\%+10\text{dig})$		
		200.00mV	0.01mV			
	V	2.0000V	0.1mV	$\pm(0.1\%+2\text{dig})$		
		20.000V	1mV			
		200.00V	10mV			
		1000.0V	0.1V	$\pm(0.15\%+5\text{dig})$		
AC Voltage (V)	mV ^[1]	20.000mV	0.001 mV	VRMS Freq range: 40Hz-1000Hz	$\pm(0.5\%+10\text{dig})$	
		200.00mV	0.01mV			
	V	2.0000V	0.1mV		$\pm(0.8\%+10\text{dig})$	
		20.000V	1mV			
		200.00V	10mV			
		750.0V	0.1V			
DC Current (A)	μA	200.00μA	0.01μA	$\pm(0.5\%+10\text{dig})$		
	mA	2.0000mA	0.1μA			
		20.000mA	1μA			
		200.00mA	10μA			
	A	20.000A ^[2]	1mA	$\pm(2.0\%+10\text{dig})$		
AC Current (A)	μA	200.00μA	0.01μA	VRMS Freq range: 40Hz-1000Hz	$\pm(0.8\%+10\text{dig})$	
	mA	2.0000mA	0.1μA			
		20.000mA	1μA			
		200.00mA	10μA			
	A	20.000A ^[2]	1mA		$\pm(2.5\%+10\text{dig})$	
Resistance (Ω)		200.00Ω	0.01Ω	$\pm(0.5\%+10\text{dig})$		
		2.0000kΩ	0.1Ω	$\pm(0.3\%+3\text{dig})$		
		20.000kΩ	1Ω	$\pm(0.3\%+1\text{dig})$		
		200.00kΩ	10Ω			
		2.0000MΩ	100Ω			
		20.000MΩ	1kΩ	$\pm(0.5\%+1\text{dig})$		

	200.00MΩ	10kΩ	±(5.0%+10dig)
Capacitance (F)	2.0000nF	0.1pF	±(3.0%+10dig)
	20.000nF	1pF	
	200.00nF	10pF	
	2.0000μF	100pF	
	20.000μF	1nF	
	200.00μF	10nF	
	2.0000mF	100nF	
	20.000mF ^[3]	1μF	
Frequency ^[4] (Hz)	200.00Hz	0.01Hz	±(0.1%+4dig)
	2.0000kHz	0.1Hz	
	20.000kHz	1Hz	
	200.00kHz	10Hz	
	2.0000MHz	0.1kHz	
	20.000MHz	1kHz	
Duty Cycle ^[5] (%)	0.1% - 99.9% (Typical: Vrms=1 V, f=1 kHz)	0.1%	±(1.2%+3dig)
	0.1% - 99.9%(≥ 1 kHz)		±(2.5%+3dig)
Temperature (°C/F)	−50 °C to 400 °C	0.1 °C	±(1.0%+3°C)
	−58 °F to 752 °F	0.1 °F	±(1.2%+6°F)

[1] The rotary switch position **mV** is only for specific models.

[2] When measuring current, for 10 A to 15 A, the measuring duration should not be over 2 minutes within 10 minutes, and in this 10 minutes, no other current should flow through except within the measuring duration; for 15 A to 20 A, the measuring duration should not be over 10 seconds within 15 minutes, and in this 15 minutes, no other current should flow through except within the measuring duration.

[3] When measuring capacitance, for the 20.00mF range, the measuring duration should be over 30 seconds.

[4] When measuring frequency, the typical waveform is Square or Sine. The signal meets the following conditions.

Frequency	Amplitude (rms)
1 Hz – 4 MHz	≥ 100 mV
4 Hz – 8 MHz	≥ 200 mV
8 Hz – 10 MHz	≥ 300 mV

[5] When measuring duty cycle, the typical waveform is Square.

Note: when measuring resistance and capacitance, the influence of the resistance reactance of the pen itself on the measured value should be considered.

OW18A/OW18B multimeter

Function		Measurement Range	Resolution	Function
DC Voltage (V)	mV [1]	60.00mV/600.0mV	0.01mV	$\pm(0.5\%+2\text{dig})$
	V	600.0mV/6.000V/60.00V/600.0V	0.1mV	
	V	1000V	1V	$\pm(0.8\%+2\text{dig})$
AC Voltage (V)	mV [1]	60.00mV/600.0mV	0.01mV	$\pm(0.8\%+3\text{dig})$
	V	600.0mV	0.1 mV	$\pm(2\%+5\text{dig})$
	V	6.000V/60.00V/600.0V	1mV	$\pm(0.8\%+3\text{dig})$
	V	750V	1V	$\pm(1\%+3\text{dig})$
DC Current (A)	μ A	600.0 μ A/6000 μ A	0.1 μ A	$\pm(0.8\%+2\text{dig})$
	mA	60.00mA/600.0mA	0.01mA	$\pm(0.8\%+2\text{dig})$
	A	20.00A [2]	0.01A	$\pm(1.2\%+3\text{dig})$
AC Current (A)	μ A	600.0 μ A/6000 μ A	0.1 μ A	$\pm(1\%+3\text{dig})$
	mA	60.00mA/600.0mA	0.01mA	$\pm(1\%+3\text{dig})$
	A	20.00A [2]	0.01A	$\pm(1.5\%+3\text{dig})$
Resistance (Ω)	600.0 Ω /6.000k Ω /60.00k Ω / 600.0k Ω /6.000M Ω		0.1 Ω	$\pm(0.8\%+2\text{dig})$
	60.00M Ω		0.01 M Ω	$\pm(2\%+3\text{dig})$
Capacitance (F)	60.00nF/600.0nF/6.000 μ F/ 60.00 μ F		0.01nF	$\pm(3\%+3\text{dig})$
	600.0 μ F/6.000mF/60.00mF [3]		0.1 μ F	$\pm(3\%+5\text{dig})$
Frequency [4] (Hz)	9.999Hz/99.99Hz/999.9Hz/ 9.999kHz/99.99kHz/999.9kHz/ 9.999MHz		0.001Hz	$\pm(0.8\%+2\text{dig})$
Duty Cycle [5] (%)	0.1% - 99.9% (Typical: Vrms=1 V, f=1 kHz)		0.1%	$\pm(1.2\%+3\text{dig})$
	0.1% - 99.9%(≥ 1 kHz)			$\pm(2.5\%+3\text{dig})$
Temperature ($^{\circ}$ C/ $^{\circ}$ F)	-50 $^{\circ}$ C to 400 $^{\circ}$ C		1 $^{\circ}$ C	$\pm(2.5\%+3\text{dig})$
	-58 $^{\circ}$ F to 752 $^{\circ}$ F		1 $^{\circ}$ F	$\pm(4.5\%+5\text{dig})$

\approx

[1] The rotary switch position **mV** is only for specific models.

[2] When measuring current, for 10 A to 15 A, the measuring duration should not be over 2 minutes within 10 minutes, and in this 10 minutes, no other current should flow through except within the measuring duration; for 15 A to 20 A, the measuring duration should not be over 10 seconds within 15 minutes, and in this 15 minutes, no other current should flow through except within the measuring duration.

[3] When measuring capacitance, for the 60.00mF range, the measuring duration should be over 30 seconds.

[4] When measuring frequency, the typical waveform is Square or Sine. The signal meets the following conditions.

Frequency	Amplitude (rms)
1 Hz – 5 MHz	≥ 700 mV

[5] When measuring duty cycle, the typical waveform is Square.

Note: when measuring resistance and capacitance, the influence of the resistance reactance of the pen itself on the measured value should be considered.

OW16 series multimeter

Function		Measurement Range	Resolution	Function
DC Voltage (V)	V	600.0mV/6.000V/60.00V/600.0V	0.1mV	$\pm(0.5\%+2\text{dig})$
		1000V	1V	$\pm(0.8\%+2\text{dig})$
AC Voltage (V)	V	600.0mV	0.1mV	$\pm(2\%+5\text{dig})$
		6.000V/60.00V/600.0V	1mV	$\pm(0.8\%+3\text{dig})$
		750V	1V	$\pm(1\%+3\text{dig})$
DC Current (A)	$\mu\text{A}^{[1]}$	600.0 μA /6000 μA	0.1 μA	$\pm(0.8\%+2\text{dig})$
	mA	60.00mA/600.0mA	0.01mA	$\pm(0.8\%+2\text{dig})$
	A	10.00A	0.01A	$\pm(1.2\%+3\text{dig})$
AC Current (A)	$\mu\text{A}^{[1]}$	600.0 μA /6000 μA	0.1 μA	$\pm(1\%+3\text{dig})$
	mA	60.00mA/600.0mA	0.01mA	$\pm(1\%+3\text{dig})$
	A	10.00A	0.01A	$\pm(1.5\%+3\text{dig})$
Resistance (Ω)	600.0 Ω /6.000k Ω /60.00k Ω / 600.0k Ω /6.000M Ω		0.1 Ω	$\pm(0.8\%+2\text{dig})$
	60.00M Ω		0.01 M Ω	$\pm(2\%+3\text{dig})$
	60.00nF/600.0nF/6.000 μF / 60.00 μF		0.01nF	$\pm(3\%+3\text{dig})$
Frequency ^[3] (Hz)		600.0 μF /6.000mF/60.00mF ^[2]	0.1 μF	$\pm(3\%+5\text{dig})$
Frequency ^[3] (Hz)		9.999Hz/99.99Hz/999.9Hz/ 9.999kHz/99.99kHz/999.9kHz/ 9.999MHz	0.001Hz	$\pm(0.8\%+2\text{dig})$
Duty Cycle ^[4] (%)	0.1% - 99.9% (Typical: Vrms=1 V, f=1 kHz)		0.1%	$\pm(1.2\%+3\text{dig})$
	0.1% - 99.9%(≥ 1 kHz)			$\pm(2.5\%+3\text{dig})$
Temperature ($^{\circ}\text{C}/^{\circ}\text{F}$)	−50 °C to 400 °C		1 °C	$\pm(2.5\%+3\text{dig})$
	−58 °F to 752 °F		1 °F	$\pm(4.5\%+5\text{dig})$

[1] The rotary switch position $\mu\text{A}\approx$ is only for specific models.

[2] When measuring capacitance, for the 60.00mF range, the measuring duration should be over 30 seconds.

[3] When measuring frequency, the typical waveform is Square or Sine. The signal meets the following conditions.

Frequency	Amplitude (rms)
1 Hz – 5 MHz	≥ 700 mV

[4] When measuring duty cycle, the typical waveform is Square.

Characteristics	Instruction	
Display	OW16.OW16B.OW18A.OW18B	5999
	OW18D.OW18E	19999
Frequency Response (Hz)	(40 - 1000) Hz	
Sample rate for digital data	3 times/second	
Bluetooth	OW16A .OW18A .OW18D	Without
	OW16B .OW18B .OW18E	✓
Auto ranging	✓	
True RMS	✓	
Diodes Test	✓	
Sleep Mode	✓	
Continuity Test	✓	
NCV function	✓	
Flashlight	✓	
Low battery indication	✓ (The "[-+]" is displayed when the battery is under the proper operation range.)	
Data Hold	✓	
Relative Measurement	✓	
LCD Backlight	✓	
Input Protection	✓	
Input Impedance	$\geq 10 M\Omega$	
Battery	OW18 series	9V battery (6F22)
	OW16 series	3 V (1.5 V × 2) AA
LCD Size	OW18series	58.5 mm * 41 mm
	OW16 series	69 mm * 52 mm
Weight (without package)	OW18series	0.32 kg
	OW16 series	0.29 kg
Dimension	OW18series	190 mm * 90 mm * 56
	OW16 series	85 mm * 185 mm * 30
Working temperature	0°C to 40°C	
Storage temperature	-10°C to 60°C	
Relative Humidity	$\leq 80\%$	
Altitude	Operating: 3,000 meters Non-operating: 15,000 meters	

Interval Period of Adjustment:

One year is recommended for the calibration interval period.

V1.3.1



7007020100054